

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071718\
 Data File : VX003429.D
 Acq On : 17 Jul 2018 15:54
 Operator : JC/MD
 Sample : J3997-16
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-H-D

Quant Time: Jul 18 01:27:39 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071618W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 16 13:15:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	261916	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	395120	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	371918	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	224311	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	174509	50.88	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	101.76%	
35) Dibromofluoromethane	5.51	113	150146	46.97	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	93.94%	
50) Toluene-d8	8.72	98	643966	54.58	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	109.16%	
62) 4-Bromofluorobenzene	11.14	95	199338	47.32	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	94.64%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.45	43	14748	10.34	ug/l	# 88
18) Methyl Acetate	2.78	43	8819	1.90	ug/l	94
20) Methylene Chloride	2.86	84	37024	12.37	ug/l	92
25) 2-Butanone	4.72	43	4067	1.82	ug/l	95
43) Isopropyl Acetate	6.48	43	129676	19.06	ug/l	94
74) N-amyl acetate	6.48	43	129676	18.77	ug/l	# 94
95) Naphthalene	13.83	128	38873	2.64	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX071718\
Data File : VX003429.D
Acq On : 17 Jul 2018 15:54
Operator : JC/MD
Sample : J3997-16
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-H-D

Quant Time: Jul 18 01:27:39 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071618W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 16 13:15:28 2018
Response via : Initial Calibration

