

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071718\
 Data File : VX003431.D
 Acq On : 17 Jul 2018 16:47
 Operator : JC/MD
 Sample : J3819-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 AU-05-071318-B

Quant Time: Jul 18 01:35:10 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071618W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 16 13:15:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	280058	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	410616	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	353365	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	209299	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	170180	46.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.80%	
35) Dibromofluoromethane	5.51	113	153766	46.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.56%	
50) Toluene-d8	8.72	98	642406	52.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.80%	
62) 4-Bromofluorobenzene	11.14	95	197250	45.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.12%	
Target Compounds						
16) Acetone	2.45	43	27554	18.06	ug/l	Qvalue # 89
18) Methyl Acetate	2.78	43	26979	5.43	ug/l	94
20) Methylene Chloride	2.86	84	18670	2.94	ug/l	95
25) 2-Butanone	4.71	43	6104	2.56	ug/l	94
95) Naphthalene	13.84	128	177359	12.92	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX071718\
Data File : VX003431.D
Acq On : 17 Jul 2018 16:47
Operator : JC/MD
Sample : J3819-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
AU-05-071318-B

Quant Time: Jul 18 01:35:10 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071618W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 16 13:15:28 2018
Response via : Initial Calibration

