

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX071718\
 Data File : VX003420.D
 Acq On : 17 Jul 2018 11:52
 Operator : JC/MD
 Sample : J4017-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-1

Quant Time: Jul 20 13:46:06 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071618W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 16 13:15:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	299628	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	422213	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	388311	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	221917	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	182041	46.39	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.78%	
35) Dibromofluoromethane	5.51	113	157563	46.12	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.24%	
50) Toluene-d8	8.72	98	673386	53.41	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.82%	
62) 4-Bromofluorobenzene	11.14	95	233162	51.80	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.60%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX071718\
Data File : VX003420.D
Acq On : 17 Jul 2018 11:52
Operator : JC/MD
Sample : J4017-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-1

Quant Time: Jul 20 13:46:06 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071618W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 16 13:15:28 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.00 ug/l

RT: 5.67 min Scan# 833

Delta R.T. 0.00 min

Lab File: VX003420.D

Acq: 17 Jul 2018 11:52

Instrument :

MSVOA_X

ClientSampleId :


MW-1

Tot Ion:168 Resp: 299628

Ion Ratio Lower Upper

168 100

99 47.7 39.3 58.9



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.67

120000

100000

80000



60000

40000

20000

0

Time--> 5.50 5.60 5.70 5.80



#33

1,2-Dichloroethane-d4

Concen: 46.39 ug/l

RT: 6.07 min Scan# 900

Delta R.T. 0.00 min

Lab File: VX003420.D


Acq: 17 Jul 2018 11:52

Tgt Ion: 65 Resp: 182041

Ion Ratio Lower Upper

65 100

67 52.4 0.0 102.6



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0

6.07

80000


60000


40000

20000

0

Time--> 5.90 6.00 6.10 6.20









#34
 1,4-Difluorobenzene
 Concen: 50.00 ug/l
 RT: 6.87 min Scan# 1030
 Delta R.T. 0.00 min
 Lab File: VX003420.D
 Acq: 17 Jul 2018 11:52

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-1

Tot Ion:114 Resp: 422213
 Ion Ratio Lower Upper
 114 100
 63 18.8 0.0 41.4
 88 15.3 0.0 30.0




Abundance Ion 113.90 (113.60 to 114.60): V
 Ion 63.00 (62.70 to 63.70): VX
 Ion 87.90 (87.60 to 88.60): VX






#35
 Dibromofluoromethane
 Concen: 46.12 ug/l
 RT: 5.51 min Scan# 807
 Delta R.T. 0.00 min
 Lab File: VX003420.D
 Acq: 17 Jul 2018 11:52

Tgt Ion:113 Resp: 157563
 Ion Ratio Lower Upper
 113 100
 111 101.2 81.4 122.0
 192 24.9 19.1 28.7



Abundance Ion 112.80 (112.50 to 113.50): V
 Ion 110.80 (110.50 to 111.50): V
 Ion 191.70 (191.40 to 192.40): V









#50
Toluene-d8
Concen: 53.41 ug/l
RT: 8.72 min Scan# 1334
Delta R.T. 0.00 min
Lab File: VX003420.D
Acq: 17 Jul 2018 11:52




Instrument :
MSVOA_X
ClientSampleId :
MW-1


Tot Ion: 98 Resp: 673386
Ion Ratio Lower Upper
98 100
100 64.6 50.9 76.3



#62
4-Bromofluorobenzene
Concen: 51.80 ug/l
RT: 11.14 min Scan# 1731
Delta R.T. 0.00 min
Lab File: VX003420.D
Acq: 17 Jul 2018 11:52

Tgt Ion: 95 Resp: 233162
Ion Ratio Lower Upper
95 100
174 94.3 0.0 187.4
176 90.1 0.0 181.0









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 10.12 min Scan# 1563
Delta R.T. 0.00 min
Lab File: VX003420.D
Acq: 17 Jul 2018 11:52

Instrument :
MSVOA_X
ClientSampleId :
MW-1

Tot Ion:117 Resp: 388311
Ion Ratio Lower Upper
117 100
82 52.7 45.0 67.4
119 31.8 25.8 38.6



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 12.08 min Scan# 1885
Delta R.T. 0.00 min
Lab File: VX003420.D
Acq: 17 Jul 2018 11:52

Tgt Ion:152 Resp: 221917
Ion Ratio Lower Upper
152 100
115 55.7 40.1 120.2
150 158.5 0.0 347.2

