

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071719\
 Data File : VX010927.D
 Acq On : 17 Jul 2019 11:50
 Operator : JC/SP
 Sample : K3834-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TW-5

Quant Time: Jul 18 03:22:19 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 17 06:57:45 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	218591	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	344241	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	315047	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	146980	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	113143	48.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.30%	
35) Dibromofluoromethane	5.50	113	107266	49.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.18%	
50) Toluene-d8	8.71	98	412465	50.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.66%	
62) 4-Bromofluorobenzene	11.13	95	141533	46.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.98%	
Target Compounds						
16) Acetone	2.45	43	1701	1.620	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX071719\
Data File : VX010927.D
Acq On : 17 Jul 2019 11:50
Operator : JC/SP
Sample : K3834-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TW-5

Quant Time: Jul 18 03:22:19 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 17 06:57:45 2019
Response via : Initial Calibration

