

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071719\
 Data File : VX010934.D
 Acq On : 17 Jul 2019 14:33
 Operator : JC/SP
 Sample : K3856-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-01-190716

Quant Time: Jul 18 03:42:16 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 17 06:57:45 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	216593	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	347068	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	320576	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	149348	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	116985	50.77	ug/l	0.00
Spiked Amount						
			Recovery	=	101.54%	
35) Dibromofluoromethane	5.50	113	107513	48.80	ug/l	0.00
Spiked Amount						
			Recovery	=	97.60%	
50) Toluene-d8	8.71	98	413242	50.02	ug/l	0.00
Spiked Amount						
			Recovery	=	100.04%	
62) 4-Bromofluorobenzene	11.13	95	145696	47.97	ug/l	0.00
Spiked Amount						
			Recovery	=	95.94%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX071719\
Data File : VX010934.D
Acq On : 17 Jul 2019 14:33
Operator : JC/SP
Sample : K3856-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-01-190716

Quant Time: Jul 18 03:42:16 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 17 06:57:45 2019
Response via : Initial Calibration

