

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX071720\
 Data File : VX017412.D
 Acq On : 17 Jul 2020 12:00
 Operator : JC/SP
 Sample : VX0717WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0717WBL01

Quant Time: Jul 18 07:19:49 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071320W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 14 02:59:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	232392	50.00	ug/l	0.01
34) 1,4-Difluorobenzene	6.84	114	416986	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	446420	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	230236	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	172969	59.30	ug/l	0.00
Spiked Amount	50.000		Recovery	=	118.60%	
35) Dibromofluoromethane	5.47	113	149502	53.22	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.44%	
50) Toluene-d8	8.70	98	524344	51.75	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.50%	
62) 4-Bromofluorobenzene	11.12	95	207964	51.67	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.34%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX071720\
Data File : VX017412.D
Acq On : 17 Jul 2020 12:00
Operator : JC/SP
Sample : VX0717WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampled :
VX0717WBL01

Quant Time: Jul 18 07:19:49 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071320W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 14 02:59:51 2020
Response via : Initial Calibration

