

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071720\
 Data File : VX017416.D
 Acq On : 17 Jul 2020 13:54
 Operator : JC/SP
 Sample : L3188-16
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TR-04-071620

Quant Time: Jul 18 07:26:16 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071320W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 14 02:59:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	256171	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	457080	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	512067	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	255871	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	178741	55.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.18%	
35) Dibromofluoromethane	5.46	113	156909	50.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.92%	
50) Toluene-d8	8.70	98	601076	54.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.24%	
62) 4-Bromofluorobenzene	11.12	95	240506	54.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.04%	
Target Compounds						
16) Acetone	2.42	43	36152	31.518	ug/l	96
18) Methyl Acetate	2.75	43	3874	1.367	ug/l	99
20) Methylene Chloride	2.84	84	37273	13.319	ug/l	99
43) Isopropyl Acetate	6.42	43	71376	10.273	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX071720\
Data File : VX017416.D
Acq On : 17 Jul 2020 13:54
Operator : JC/SP
Sample : L3188-16
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TR-04-071620

Quant Time: Jul 18 07:26:16 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071320W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 14 02:59:51 2020
Response via : Initial Calibration

