

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071720\
 Data File : VX017419.D
 Acq On : 17 Jul 2020 15:02
 Operator : JC/SP
 Sample : PB130263ZHE#11
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB130263ZHE#11

Quant Time: Jul 18 07:35:59 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071320W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 14 02:59:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	243517	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	429443	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	466121	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	245026	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	175882	57.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	115.10%	
35) Dibromofluoromethane	5.47	113	152019	52.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.10%	
50) Toluene-d8	8.70	98	552552	52.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.90%	
62) 4-Bromofluorobenzene	11.12	95	224316	54.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.24%	
Target Compounds						
16) Acetone	2.42	43	6974	6.396	ug/l	94
20) Methylene Chloride	2.84	84	25531	9.597	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX071720\
Data File : VX017419.D
Acq On : 17 Jul 2020 15:02
Operator : JC/SP
Sample : PB130263ZHE#11
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB130263ZHE#11

Quant Time: Jul 18 07:35:59 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071320W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 14 02:59:51 2020
Response via : Initial Calibration

