

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071724\
 Data File : VX042357.D
 Acq On : 17 Jul 2024 14:43
 Operator : JC/MD
 Sample : P3282-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IDW-T1-01

Quant Time: Jul 18 00:57:51 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 23:51:58 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	145821	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	226902	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	201307	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	104528	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	88146	52.421	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.840%
35) Dibromofluoromethane	5.379	113	87615	54.420	ug/l	-0.01
Spiked Amount	50.000	Range	75 - 124	Recovery	=	108.840%
50) Toluene-d8	8.647	98	256663	45.379	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	90.760%#
62) 4-Bromofluorobenzene	11.079	95	97932	46.971	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.940%
Target Compounds						
						Qvalue
16) Acetone	2.380	43	6227	8.434	ug/l	# 84
52) Toluene	8.720	92	5982	1.629	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071724\
Data File : VX042357.D
Acq On : 17 Jul 2024 14:43
Operator : JC/MD
Sample : P3282-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IDW-T1-01

Quant Time: Jul 18 00:57:51 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 21 23:51:58 2024
Response via : Initial Calibration

