



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071724\
Data File : VX042351.D
Acq On : 17 Jul 2024 09:49
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
Title : SW846 8260
Last Update : Fri Jun 21 23:51:58 2024



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	15.2	7877	PASS
75	95	30	60	45.4	23595	PASS
95	95	100	100	100.0	51944	PASS
96	95	5	9	6.6	3417	PASS
173	174	0.00	2	0.3	130	PASS
174	95	50	100	96.3	50045	PASS
175	174	5	9	7.8	3919	PASS
176	174	95	101	98.2	49155	PASS
177	176	5	9	7.1	3486	PASS