

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071818\
 Data File : VX003455.D
 Acq On : 18 Jul 2018 18:41
 Operator : JC/MD
 Sample : PB111206ZHE#01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB111206ZHE#01

Quant Time: Jul 19 03:20:14 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071618W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 16 13:15:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	297725	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	443531	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	393716	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	213248	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	182682	46.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.70%	
35) Dibromofluoromethane	5.51	113	163195	45.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.96%	
50) Toluene-d8	8.72	98	594569	44.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.80%	
62) 4-Bromofluorobenzene	11.14	95	205425	43.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.90%	
Target Compounds						
16) Acetone	2.45	43	19070	11.76	ug/l #	88
18) Methyl Acetate	2.78	43	7884	1.49	ug/l	93
20) Methylene Chloride	2.86	84	6716	Below Cal		93
43) Isopropyl Acetate	6.47	43	152767	20.00	ug/l	94
74) N-amyl acetate	6.47	43	152767	23.26	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX071818\
Data File : VX003455.D
Acq On : 18 Jul 2018 18:41
Operator : JC/MD
Sample : PB111206ZHE#01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB111206ZHE#01

Quant Time: Jul 19 03:20:14 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071618W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 16 13:15:28 2018
Response via : Initial Calibration

