


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX071818\
 Data File : VX003438.D
 Acq On : 18 Jul 2018 09:07
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071618W.M
 Title : SW846 8260
 Last Update : Mon Jul 16 13:15:28 2018



AutoFind: Scans 1730, 1731, 1732; Background Corrected with Scan 1723

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.0	4944	PASS
75	95	30	60	47.1	14551	PASS
95	95	100	100	100.0	30898	PASS
96	95	5	9	7.1	2182	PASS
173	174	0.00	2	1.1	349	PASS
174	95	50	100	98.2	30354	PASS
175	174	5	9	7.8	2379	PASS
176	174	95	101	96.4	29266	PASS
177	176	5	9	6.8	1995	PASS