

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071819\
 Data File : VX010949.D
 Acq On : 18 Jul 2019 11:29
 Operator : JC/SP
 Sample : VX0718MBL01
 Misc : 5.0g/10mL/100uL/5mL/MSVOA_X/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0718MBL01

Quant Time: Jul 18 14:14:20 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 16 12:41:18 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	206528	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	323021	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	296858	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	133532	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	110805	50.43	ug/l	0.00
Spiked Amount			Recovery	=	100.86%	
35) Dibromofluoromethane	5.50	113	101452	49.48	ug/l	0.00
Spiked Amount			Recovery	=	98.96%	
50) Toluene-d8	8.71	98	387534	50.40	ug/l	0.00
Spiked Amount			Recovery	=	100.80%	
62) 4-Bromofluorobenzene	11.13	95	131921	46.67	ug/l	0.00
Spiked Amount			Recovery	=	93.34%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX071819\
Data File : VX010949.D
Acq On : 18 Jul 2019 11:29
Operator : JC/SP
Sample : VX0718MBL01
Misc : 5.0µ/10mL/100uL/5mL/MSVOA_X/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0718MBL01

Quant Time: Jul 18 14:14:20 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 16 12:41:18 2019
Response via : Initial Calibration

