

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071819\
 Data File : VX010957.D
 Acq On : 18 Jul 2019 14:36
 Operator : JC/SP
 Sample : K3857-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 36916

Quant Time: Jul 19 02:09:55 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 17 06:57:45 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	210876	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	336667	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	311530	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	147200	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	112378	50.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.18%	
35) Dibromofluoromethane	5.50	113	102430	47.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.86%	
50) Toluene-d8	8.71	98	402237	50.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.38%	
62) 4-Bromofluorobenzene	11.13	95	140216	47.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.20%	
Target Compounds						
16) Acetone	2.45	43	4633	4.573	ug/l	93
29) Tetrahydrofuran	5.13	42	12155	13.193	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX071819\
Data File : VX010957.D
Acq On : 18 Jul 2019 14:36
Operator : JC/SP
Sample : K3857-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
36916

Quant Time: Jul 19 02:09:55 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 17 06:57:45 2019
Response via : Initial Calibration

