

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071819\
 Data File : VX010948.D
 Acq On : 18 Jul 2019 11:06
 Operator : JC/SP
 Sample : VX0718WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0718WBL01

Quant Time: Jul 19 01:45:29 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 17 06:57:45 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	215399	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	339361	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	301975	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	131968	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	112343	49.03	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.06%	
35) Dibromofluoromethane	5.49	113	107128	49.73	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.46%	
50) Toluene-d8	8.71	98	400528	49.58	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.16%	
62) 4-Bromofluorobenzene	11.13	95	132944	44.77	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.54%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX071819\
Data File : VX010948.D
Acq On : 18 Jul 2019 11:06
Operator : JC/SP
Sample : VX0718WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0718WBL01

Quant Time: Jul 19 01:45:29 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 17 06:57:45 2019
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.66 min Scan# 739
Delta R.T. -0.00 min
Lab File: VX010948.D
Acq: 18 Jul 2019 11:06




Instrument :
MSVOA_X
ClientSampleId :
VX0718WBL01


Tot Ion:168 Resp: 215399
Ion Ratio Lower Upper
168 100
99 48.3 37.6 56.4



#33
1,2-Dichloroethane-d4
Concen: 49.026 ug/l
RT: 6.06 min Scan# 804
Delta R.T. -0.01 min
Lab File: VX010948.D
Acq: 18 Jul 2019 11:06

Tgt Ion: 65 Resp: 112343
Ion Ratio Lower Upper
65 100
67 53.8 0.0 107.8





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.85 min Scan# 935

Delta R.T. -0.01 min

Lab File: VX010948.D

Acq: 18 Jul 2019 11:06

Instrument :

MSVOA_X

ClientSampleId :

VX0718WBL01


Tot Ion:114 Resp: 339361

Ion Ratio Lower Upper

114 100

63 18.2 0.0 35.6




88 14.4 0.0 29.8



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 49.735 ug/l

RT: 5.49 min Scan# 712

Delta R.T. -0.01 min

Lab File: VX010948.D

Acq: 18 Jul 2019 11:06


Tgt Ion:113 Resp: 107128

Ion Ratio Lower Upper

113 100

111 101.9 82.3 123.5



192 21.4 17.4 26.2




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V









#50
Toluene-d8
Concen: 49.577 ug/l
RT: 8.71 min Scan# 1240
Delta R.T. -0.00 min
Lab File: VX010948.D
Acq: 18 Jul 2019 11:06




Instrument :
MSVOA_X
ClientSampleId :
VX0718WBL01


Tot Ion: 98 Resp: 400528
Ion Ratio Lower Upper
98 100
100 64.2 50.6 75.8



#62
4-Bromofluorobenzene
Concen: 44.769 ug/l
RT: 11.13 min Scan# 1637
Delta R.T. -0.00 min
Lab File: VX010948.D
Acq: 18 Jul 2019 11:06

Tgt Ion: 95 Resp: 132944
Ion Ratio Lower Upper
95 100
174 92.2 0.0 180.6
176 88.5 0.0 173.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.11 min Scan# 1469
Delta R.T. -0.00 min
Lab File: VX010948.D
Acq: 18 Jul 2019 11:06

Instrument :
MSVOA_X
ClientSampleId :
VX0718WBL01

Tot Ion:117 Resp: 301975
Ion Ratio Lower Upper
117 100
82 50.7 42.6 63.8
119 32.5 25.9 38.9




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): VX0
Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.08 min Scan# 1792
Delta R.T. -0.00 min
Lab File: VX010948.D
Acq: 18 Jul 2019 11:06

Tgt Ion:152 Resp: 131968
Ion Ratio Lower Upper
152 100
115 55.1 39.7 119.1
150 157.0 0.0 345.6



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V

