

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071822\
 Data File : VX030104.D
 Acq On : 18 Jul 2022 12:38
 Operator : JC/MD
 Sample : N3755-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-01-071522

Quant Time: Jul 19 05:40:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X071222W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 13 09:05:23 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	175360	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	317698	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	279154	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	119571	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	110808	51.914	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	103.820%
35) Dibromofluoromethane	5.391	113	96668	47.630	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.260%
50) Toluene-d8	8.653	98	366905	48.067	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.140%
62) 4-Bromofluorobenzene	11.085	95	118157	44.606	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.220%
Target Compounds						
16) Acetone	2.386	43	2609	2.594	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071822\
Data File : VX030104.D
Acq On : 18 Jul 2022 12:38
Operator : JC/MD
Sample : N3755-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-01-071522

Quant Time: Jul 19 05:40:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X071222W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 13 09:05:23 2022
Response via : Initial Calibration

