

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX071822\  
 Data File : VX030105.D  
 Acq On : 18 Jul 2022 13:02  
 Operator : JC/MD  
 Sample : N3755-04  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 9 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 EB-01-071522

Quant Time: Jul 19 05:40:43 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X071222W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Jul 13 09:05:23 2022  
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
-----						
Internal Standards						
1) Pentafluorobenzene	5.562	168	173953	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	311189	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	274221	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	113113	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	108559	51.272	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	102.540%
35) Dibromofluoromethane	5.397	113	94708	47.640	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.280%
50) Toluene-d8	8.653	98	358508	47.950	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.900%
62) 4-Bromofluorobenzene	11.085	95	120107	46.290	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.580%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	20233	20.280	ug/l	97
29) Tetrahydrofuran	5.025	42	1504	1.304	ug/l	89
-----						

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX071822\  
 Data File : VX030105.D  
 Acq On : 18 Jul 2022 13:02  
 Operator : JC/MD  
 Sample : N3755-04  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 9 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 EB-01-071522

Quant Time: Jul 19 05:40:43 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X071222W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Jul 13 09:05:23 2022  
 Response via : Initial Calibration

