

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071822\
 Data File : VX030100.D
 Acq On : 18 Jul 2022 10:53
 Operator : JC/MD
 Sample : VX0718WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0718WBL01

Quant Time: Jul 19 05:39:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X071222W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 13 09:05:23 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	175843	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	311630	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	262818	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	100352	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	108549	50.716	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	101.440%
35) Dibromofluoromethane	5.391	113	94351	47.393	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.780%
50) Toluene-d8	8.653	98	350401	46.799	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.600%
62) 4-Bromofluorobenzene	11.079	95	110688	42.600	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	85.200%


Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071822\
Data File : VX030100.D
Acq On : 18 Jul 2022 10:53
Operator : JC/MD
Sample : VX0718WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0718WBL01

Quant Time: Jul 19 05:39:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X071222W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 13 09:05:23 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 713
Delta R.T. -0.006 min
Lab File: VX030100.D
Acq: 18 Jul 2022 10:53




Instrument :
MSVOA_X
ClientSampleId :
VX0718WBL01


Tgt Ion:168 Resp: 175843
Ion Ratio Lower Upper
168 100
99 53.3 44.0 66.0



#33
1,2-Dichloroethane-d4
Concen: 50.716 ug/l
RT: 5.958 min Scan# 799
Delta R.T. 0.000 min
Lab File: VX030100.D
Acq: 18 Jul 2022 10:53

Tgt Ion: 65 Resp: 108549
Ion Ratio Lower Upper
65 100
67 53.6 0.0 109.6









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.763 min Scan# 911
Delta R.T. -0.000 min
Lab File: VX030100.D
Acq: 18 Jul 2022 10:53




Instrument :
MSVOA_X
ClientSampleId :
VX0718WBL01


Tgt Ion:114 Resp: 311630
Ion Ratio Lower Upper
114 100
63 20.2 0.0 38.4
88 15.1 0.0 31.8



#35
Dibromofluoromethane
Concen: 47.393 ug/l
RT: 5.391 min Scan# 706
Delta R.T. 0.000 min
Lab File: VX030100.D
Acq: 18 Jul 2022 10:53

Tgt Ion:113 Resp: 94351
Ion Ratio Lower Upper
113 100
111 100.8 83.2 124.8
192 18.3 14.8 22.2









#50
Toluene-d8
Concen: 46.799 ug/l
RT: 8.653 min Scan# 1241
Delta R.T. -0.000 min
Lab File: VX030100.D
Acq: 18 Jul 2022 10:53




Instrument :
MSVOA_X
ClientSampleId :
VX0718WBL01


Tgt Ion: 98 Resp: 350401
Ion Ratio Lower Upper
98 100
100 65.8 52.3 78.5



#62
4-Bromofluorobenzene
Concen: 42.600 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX030100.D
Acq: 18 Jul 2022 10:53

Tgt Ion: 95 Resp: 110688
Ion Ratio Lower Upper
95 100
174 74.5 0.0 149.0
176 71.5 0.0 146.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX030100.D
Acq: 18 Jul 2022 10:53

Instrument :
MSVOA_X
ClientSampleId :
VX0718WBL01

Tgt Ion:117 Resp: 262818
Ion Ratio Lower Upper
117 100
82 53.1 42.7 64.1
119 32.8 25.4 38.2



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VX030100.D
Acq: 18 Jul 2022 10:53

Tgt Ion:152 Resp: 100352
Ion Ratio Lower Upper
152 100
115 60.4 40.5 121.5
150 159.9 0.0 347.0

