

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071824\
 Data File : VX042363.D
 Acq On : 18 Jul 2024 12:49
 Operator : JC/MD
 Sample : VX0718MBL01
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Jul 19 05:29:39 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML070324WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jul 19 05:29:00 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.757	114	264685	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.049	117	243599	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.018	152	116086	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	53142	34.494	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	68.980%
7) Chloroethane-d5	1.666	69	47686	49.159	ug/L	0.02
Spiked Amount	50.000	Range	70 - 130	Recovery	=	98.320%
11) 1,1-Dichloroethene-d2	2.307	65	25989	38.544	ug/L	0.01
Spiked Amount	50.000	Range	60 - 125	Recovery	=	77.080%
21) 2-Butanone-d5	4.453	46	108422	97.656	ug/L	-0.01
Spiked Amount	100.000	Range	40 - 130	Recovery	=	97.660%
24) Chloroform-d	5.056	84	128514	41.487	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	82.980%
26) 1,2-Dichloroethane-d4	5.952	65	78361	42.858	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	85.720%
32) Benzene-d6	5.971	84	261576	41.543	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	83.080%
36) 1,2-Dichloropropane-d6	7.306	67	80631	42.552	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	85.100%
41) Toluene-d8	8.647	98	236160	40.248	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	80.500%
43) trans-1,3-Dichloroprop...	8.946	79	34097	41.635	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	83.280%
47) 2-Hexanone-d5	9.385	63	78706	91.506	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	91.510%
56) 1,1,2,2-Tetrachloroeth...	11.189	84	121399	43.335	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	86.680%
66) 1,2-Dichlorobenzene-d4	12.317	152	100701	44.661	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	89.320%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071824\
Data File : VX042363.D
Acq On : 18 Jul 2024 12:49
Operator : JC/MD
Sample : VX0718MBL01
Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :

Quant Time: Jul 19 05:29:39 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML070324WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jul 19 05:29:00 2024
Response via : Initial Calibration

