

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071824\
 Data File : VX042382.D
 Acq On : 18 Jul 2024 20:05
 Operator : JC/MD
 Sample : P3238-15ME
 Misc : 4.93g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 A4BZ5ME

Quant Time: Jul 19 05:34:14 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML070324WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jul 19 05:29:00 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.757	114	235342	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	213204	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	107396	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	56510	41.253	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	82.500%
7) Chloroethane-d5	1.648	69	38090	44.163	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	88.320%
11) 1,1-Dichloroethene-d2	2.282	65	28301	47.206	ug/L	-0.01
Spiked Amount	50.000	Range	60 - 125	Recovery	=	94.420%
21) 2-Butanone-d5	4.483	46	127960	129.624	ug/L	0.02
Spiked Amount	100.000	Range	40 - 130	Recovery	=	129.620%
24) Chloroform-d	5.056	84	153541	55.747	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	111.500%
26) 1,2-Dichloroethane-d4	5.952	65	97900	60.221	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	120.440%
32) Benzene-d6	5.964	84	317662	57.643	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	115.280%
36) 1,2-Dichloropropane-d6	7.306	67	100676	60.705	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	121.420%#
41) Toluene-d8	8.647	98	286406	55.770	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	111.540%
43) trans-1,3-Dichloroprop...	8.952	79	38929	54.312	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	108.620%
47) 2-Hexanone-d5	9.391	63	102591	136.280	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	136.280%#
56) 1,1,2,2-Tetrachloroeth...	11.195	84	155630	63.475	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	126.940%#
66) 1,2-Dichlorobenzene-d4	12.317	152	132562	63.549	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	127.100%#
Target Compounds						Qvalue
15) Methyl Acetate	2.715	43	11256	6.578	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071824\
Data File : VX042382.D
Acq On : 18 Jul 2024 20:05
Operator : JC/MD
Sample : P3238-15ME
Misc : 4.93g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
A4BZ5ME

Quant Time: Jul 19 05:34:14 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML070324WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jul 19 05:29:00 2024
Response via : Initial Calibration

