

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071918\
 Data File : VX003481.D
 Acq On : 19 Jul 2018 23:40
 Operator : JC/SP
 Sample : J3170-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT101D1-071018

Quant Time: Jul 20 11:42:15 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071918W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 19 17:09:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	300703	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	442801	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	417347	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	227280	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	181785	51.07	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	102.14%	
35) Dibromofluoromethane	5.50	113	172331	51.10	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	102.20%	
50) Toluene-d8	8.72	98	629021	49.62	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	99.24%	
62) 4-Bromofluorobenzene	11.14	95	215932	46.68	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	93.36%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) Dichlorodifluoromethane	1.20	85	3823	0.89	ug/l	95
9) 1,1,2-Trichlorotrifluoroet	2.39	101	37737	11.31	ug/l	96
12) 1,1-Dichloroethene	2.37	96	18306	5.96	ug/l	86
16) Acetone	2.45	43	3387	2.52	ug/l	99
24) 1,1-Dichloroethane	3.70	63	7060	1.18	ug/l	99
27) cis-1,2-Dichloroethene	4.60	96	8474	2.23	ug/l #	11
30) Chloroform	5.22	83	6394	1.09	ug/l	96
32) 1,1,1-Trichloroethane	5.50	97	2921	0.58	ug/l #	49
38) Carbon Tetrachloride	5.79	117	6223	1.33	ug/l	96
44) Trichloroethene	7.22	130	687788	159.78	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX071918\
Data File : VX003481.D
Acq On : 19 Jul 2018 23:40
Operator : JC/SP
Sample : J3170-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT101D1-071018

Quant Time: Jul 20 11:42:15 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071918W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 19 17:09:44 2018
Response via : Initial Calibration

