

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071918\
 Data File : VX003486.D
 Acq On : 20 Jul 2018 01:53
 Operator : JC/SP
 Sample : J3170-07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE125D1-071118

Quant Time: Jul 20 04:51:26 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071918W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 19 17:09:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.68	168	292778	50.00	ug/l	0.01
34) 1,4-Difluorobenzene	6.87	114	426947	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	403657	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	219222	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	180196	51.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.98%	
35) Dibromofluoromethane	5.51	113	169260	52.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.10%	
50) Toluene-d8	8.72	98	616115	50.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.82%	
62) 4-Bromofluorobenzene	11.14	95	208935	46.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.68%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
9) 1,1,2-Trichlorotrifluoroet	2.39	101	20065	6.18	ug/l	99
12) 1,1-Dichloroethene	2.38	96	5642	1.89	ug/l	82
16) Acetone	2.45	43	3452	2.64	ug/l #	88
19) Methyl tert-butyl Ether	3.20	73	5011	0.51	ug/l	94
24) 1,1-Dichloroethane	3.70	63	9987	1.71	ug/l	99
27) cis-1,2-Dichloroethene	4.60	96	17166	4.63	ug/l	65
30) Chloroform	5.22	83	5187	0.91	ug/l	91
38) Carbon Tetrachloride	5.78	117	1139	0.25	ug/l #	79
44) Trichloroethene	7.22	130	570527	137.46	ug/l	100
49) 1,4-Dioxane	7.76	88	1916	24.17	ug/l #	88
64) Tetrachloroethene	9.34	164	12001	2.44	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX071918\
Data File : VX003486.D
Acq On : 20 Jul 2018 01:53
Operator : JC/SP
Sample : J3170-07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE125D1-071118

Quant Time: Jul 20 04:51:26 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071918W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 19 17:09:44 2018
Response via : Initial Calibration

