

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071918\
 Data File : VX003493.D
 Acq On : 20 Jul 2018 04:58
 Operator : JC/SP
 Sample : J3170-14
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GW02-071118

Manual Integrations
 APPROVED

sam
 7/20/2018 5:47:38 PM

Quant Time: Jul 26 03:45:30 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071918W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 19 17:09:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.68	168	300519	50.00	ug/l	0.01
34) 1,4-Difluorobenzene	6.87	114	421756	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	405698	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	225899	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	186426	52.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.82%	
35) Dibromofluoromethane	5.50	113	176951	55.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.16%	
50) Toluene-d8	8.72	98	624832	51.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.50%	
62) 4-Bromofluorobenzene	11.14	95	210874	47.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.72%	
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluoroet	2.39	101	48238	14.470	ug/l	95
12) 1,1-Dichloroethene	2.37	96	28610	9.321	ug/l	89
16) Acetone	2.45	43	3274	2.437	ug/l	92
24) 1,1-Dichloroethane	3.70	63	11918	1.987	ug/l	98
27) cis-1,2-Dichloroethene	4.60	96	13217	3.476	ug/l	38
30) Chloroform	5.21	83	4287	0.732	ug/l #	83
32) 1,1,1-Trichloroethane	5.49	97	3787	0.755	ug/l #	50
38) Carbon Tetrachloride	5.79	117	2207	0.494	ug/l	95
44) Trichloroethene	7.22	130	2724795	664.571	ug/l	98
49) 1,4-Dioxane	7.76	88	2306	29.443	ug/l	99
55) 1,1,2-Trichloroethane	9.22	97	3246m	0.943	ug/l	
64) Tetrachloroethene	9.34	164	2180	0.441	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX071918\
Data File : VX003493.D
Acq On : 20 Jul 2018 04:58
Operator : JC/SP
Sample : J3170-14
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GW02-071118

Manual Integrations
APPROVED

sam
7/20/2018 5:47:38 PM

Quant Time: Jul 26 03:45:30 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071918W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 19 17:09:44 2018
Response via : Initial Calibration

