

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071918\
 Data File : VX003494.D
 Acq On : 20 Jul 2018 05:25
 Operator : JC/SP
 Sample : J3965-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LF1-0034

Quant Time: Jul 20 06:33:42 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071918W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 19 17:09:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	314906	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	435450	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	390446	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	210856	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	178399	47.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.72%	
35) Dibromofluoromethane	5.51	113	179574	54.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.28%	
50) Toluene-d8	8.72	98	593124	47.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.16%	
62) 4-Bromofluorobenzene	11.14	95	203532	44.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.48%	
Target Compounds						
16) Acetone	2.45	43	5592	3.97	ug/l	# 83
44) Trichloroethene	7.21	130	11781	2.78	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX071918\
Data File : VX003494.D
Acq On : 20 Jul 2018 05:25
Operator : JC/SP
Sample : J3965-09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampled :
LF1-0034

Quant Time: Jul 20 06:33:42 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071918W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 19 17:09:44 2018
Response via : Initial Calibration

