

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071921\
 Data File : VX023332.D
 Acq On : 19 Jul 2021 10:24
 Operator : JC/MD
 Sample : VX0719MBL01
 Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0719MBL01

Quant Time: Jul 20 05:22:12 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	217205	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	351308	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	295762	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	117794	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	148623	50.912	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.820%
35) Dibromofluoromethane	5.391	113	114938	50.665	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.320%
50) Toluene-d8	8.653	98	413346	48.845	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.680%
62) 4-Bromofluorobenzene	11.085	95	140581	43.031	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.060%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071921\
Data File : VX023332.D
Acq On : 19 Jul 2021 10:24
Operator : JC/MD
Sample : VX0719MBL01
Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0719MBL01

Quant Time: Jul 20 05:22:12 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

