

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071921\
 Data File : VX023336.D
 Acq On : 19 Jul 2021 13:00
 Operator : JC/MD
 Sample : PB137749TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB137749TB

Quant Time: Jul 20 05:23:52 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	221112	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	361825	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	335037	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	144772	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	153675	51.712	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.420%
35) Dibromofluoromethane	5.397	113	117964	50.487	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.980%
50) Toluene-d8	8.653	98	440747	50.569	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.140%
62) 4-Bromofluorobenzene	11.085	95	161909	48.119	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.240%
Target Compounds						
						Qvalue
16) Acetone	2.392	43	39767	33.824	ug/l	92
20) Methylene Chloride	2.794	84	21186	8.370	ug/l	94
43) Isopropyl Acetate	6.355	43	31729	5.421	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071921\
Data File : VX023336.D
Acq On : 19 Jul 2021 13:00
Operator : JC/MD
Sample : PB137749TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB137749TB

Quant Time: Jul 20 05:23:52 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

