

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071921\
 Data File : VX023337.D
 Acq On : 19 Jul 2021 13:22
 Operator : JC/MD
 Sample : PB137749ZHE#01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB137749ZHE#01

Quant Time: Jul 20 05:24:09 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	217789	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	355328	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	326597	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	133623	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	154845	52.901	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 105.800%			
35) Dibromofluoromethane	5.397	113	118777	51.765	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 103.520%			
50) Toluene-d8	8.653	98	433418	50.637	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 101.280%			
62) 4-Bromofluorobenzene	11.085	95	155825	47.158	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 94.320%			
Target Compounds					Qvalue	
16) Acetone	2.386	43	41369	35.723	ug/l	91
20) Methylene Chloride	2.794	84	22057	8.847	ug/l #	89
43) Isopropyl Acetate	6.355	43	31941	5.557	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071921\
Data File : VX023337.D
Acq On : 19 Jul 2021 13:22
Operator : JC/MD
Sample : PB137749ZHE#01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB137749ZHE#01

Quant Time: Jul 20 05:24:09 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

