

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071921\
 Data File : VX023338.D
 Acq On : 19 Jul 2021 13:46
 Operator : JC/MD
 Sample : PB137749ZHE#02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB137749ZHE#02

Quant Time: Jul 20 05:24:26 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	214119	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	343832	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	320202	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	136408	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	151111	52.510	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 105.020%	
35) Dibromofluoromethane	5.397	113	114691	51.655	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.320%	
50) Toluene-d8	8.653	98	421019	50.833	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 101.660%	
62) 4-Bromofluorobenzene	11.085	95	154744	48.396	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 96.800%	
Target Compounds						Qvalue
16) Acetone	2.386	43	40635	35.691	ug/l	91
20) Methylene Chloride	2.794	84	22134	9.030	ug/l	97
43) Isopropyl Acetate	6.348	43	30572	5.497	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071921\
Data File : VX023338.D
Acq On : 19 Jul 2021 13:46
Operator : JC/MD
Sample : PB137749ZHE#02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB137749ZHE#02

Quant Time: Jul 20 05:24:26 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

