

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071921\
 Data File : VX023339.D
 Acq On : 19 Jul 2021 14:10
 Operator : JC/MD
 Sample : PB137749ZHE#03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB137749ZHE#03

Quant Time: Jul 20 05:24:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	212999	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	348371	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	321867	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	140708	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	149289	52.150	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.300%
35) Dibromofluoromethane	5.397	113	114930	51.088	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.180%
50) Toluene-d8	8.653	98	418690	49.894	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.780%
62) 4-Bromofluorobenzene	11.085	95	157954	48.757	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.520%
Target Compounds						Qvalue
16) Acetone	2.386	43	39435	34.819	ug/l	94
20) Methylene Chloride	2.794	84	21981	9.014	ug/l	95
43) Isopropyl Acetate	6.348	43	31561	5.601	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071921\
Data File : VX023339.D
Acq On : 19 Jul 2021 14:10
Operator : JC/MD
Sample : PB137749ZHE#03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB137749ZHE#03

Quant Time: Jul 20 05:24:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

