

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071921\
 Data File : VX023341.D
 Acq On : 19 Jul 2021 14:55
 Operator : JC/MD
 Sample : PB137749ZHE#05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB137749ZHE#05

Quant Time: Jul 20 05:25:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	212012	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	348035	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	311923	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	125526	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	149808	52.575	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 105.140%			
35) Dibromofluoromethane	5.397	113	114919	51.133	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 102.260%			
50) Toluene-d8	8.653	98	412686	49.226	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 98.460%			
62) 4-Bromofluorobenzene	11.085	95	148147	45.773	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 91.540%			
Target Compounds					Qvalue	
16) Acetone	2.386	43	21811	19.348	ug/l	90
20) Methylene Chloride	2.794	84	15227	6.274	ug/l	94
43) Isopropyl Acetate	6.355	43	30042	5.336	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071921\
Data File : VX023341.D
Acq On : 19 Jul 2021 14:55
Operator : JC/MD
Sample : PB137749ZHE#05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB137749ZHE#05

Quant Time: Jul 20 05:25:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

