

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071921\
 Data File : VX023342.D
 Acq On : 19 Jul 2021 15:18
 Operator : JC/MD
 Sample : PB137749ZHE#06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB137749ZHE#06

Quant Time: Jul 20 05:25:34 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	213771	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	349476	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	321903	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	136621	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	151414	52.701	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	105.400%
35) Dibromofluoromethane	5.398	113	114635	50.796	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.600%
50) Toluene-d8	8.653	98	426425	50.655	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.300%
62) 4-Bromofluorobenzene	11.086	95	156604	48.187	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.380%
Target Compounds						
						Qvalue
16) Acetone	2.392	43	42637	37.510	ug/l	# 87
20) Methylene Chloride	2.794	84	22650	9.255	ug/l	94
43) Isopropyl Acetate	6.355	43	32150	5.687	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071921\
Data File : VX023342.D
Acq On : 19 Jul 2021 15:18
Operator : JC/MD
Sample : PB137749ZHE#06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB137749ZHE#06

Quant Time: Jul 20 05:25:34 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

