

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071921\
 Data File : VX023346.D
 Acq On : 19 Jul 2021 17:14
 Operator : JC/MD
 Sample : PB137749ZHE#11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB137749ZHE#11

Quant Time: Jul 20 05:26:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	217202	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	357156	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	328489	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	140992	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	150594	51.588	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 103.180%	
35) Dibromofluoromethane	5.397	113	117359	50.885	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.760%	
50) Toluene-d8	8.653	98	431832	50.194	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.380%	
62) 4-Bromofluorobenzene	11.085	95	147225	44.327	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 88.660%	
Target Compounds						
						Qvalue
16) Acetone	2.386	43	22157	19.185	ug/l	94
20) Methylene Chloride	2.800	84	15907	6.397	ug/l	92
43) Isopropyl Acetate	6.348	43	28289	4.897	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071921\
Data File : VX023346.D
Acq On : 19 Jul 2021 17:14
Operator : JC/MD
Sample : PB137749ZHE#11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB137749ZHE#11

Quant Time: Jul 20 05:26:42 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

