

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071921\
 Data File : VX023347.D
 Acq On : 19 Jul 2021 17:37
 Operator : JC/MD
 Sample : PB137749ZHE#12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB137749ZHE#12

Quant Time: Jul 20 05:27:01 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	215017	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	357780	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	326416	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	130870	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	156037	53.995	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	108.000%
35) Dibromofluoromethane	5.397	113	117300	50.771	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.540%
50) Toluene-d8	8.653	98	426033	49.433	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.860%
62) 4-Bromofluorobenzene	11.085	95	154130	46.325	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.640%
Target Compounds						Qvalue
16) Acetone	2.392	43	18867	16.502	ug/l	89
20) Methylene Chloride	2.800	84	14429	5.862	ug/l	94
43) Isopropyl Acetate	6.354	43	29625	5.119	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071921\
Data File : VX023347.D
Acq On : 19 Jul 2021 17:37
Operator : JC/MD
Sample : PB137749ZHE#12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB137749ZHE#12

Quant Time: Jul 20 05:27:01 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

