

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071921\
 Data File : VX023351.D
 Acq On : 19 Jul 2021 19:12
 Operator : JC/MD
 Sample : PB137749ZHE#16
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB137749ZHE#16

Quant Time: Jul 20 05:28:07 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	216436	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	355611	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	326486	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	142093	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	153924	52.915	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 105.820%			
35) Dibromofluoromethane	5.397	113	115106	50.125	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 100.240%			
50) Toluene-d8	8.653	98	426384	49.776	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 99.560%			
62) 4-Bromofluorobenzene	11.085	95	159319	48.177	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 96.360%			
Target Compounds					Qvalue	
16) Acetone	2.392	43	18852	16.381	ug/l #	82
20) Methylene Chloride	2.800	84	14819	5.981	ug/l	92
43) Isopropyl Acetate	6.360	43	29120	5.062	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071921\
Data File : VX023351.D
Acq On : 19 Jul 2021 19:12
Operator : JC/MD
Sample : PB137749ZHE#16
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB137749ZHE#16

Quant Time: Jul 20 05:28:07 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

