

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071921\
 Data File : VX023352.D
 Acq On : 19 Jul 2021 19:36
 Operator : JC/MD
 Sample : PB137749ZHE#17
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB137749ZHE#17

Quant Time: Jul 20 05:28:23 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	210604	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	346737	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	322831	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	139485	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.971	65	150899	53.311	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 106.620%	
35) Dibromofluoromethane	5.397	113	113493	50.687	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.380%	
50) Toluene-d8	8.653	98	422980	50.642	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 101.280%	
62) 4-Bromofluorobenzene	11.085	95	156641	48.579	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 97.160%	
Target Compounds						
						Qvalue
16) Acetone	2.386	43	17729	15.832	ug/l	91
20) Methylene Chloride	2.800	84	15621	6.479	ug/l	95
43) Isopropyl Acetate	6.355	43	28871	5.147	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071921\
Data File : VX023352.D
Acq On : 19 Jul 2021 19:36
Operator : JC/MD
Sample : PB137749ZHE#17
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB137749ZHE#17

Quant Time: Jul 20 05:28:23 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

