

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071921\
 Data File : VX023353.D
 Acq On : 19 Jul 2021 19:59
 Operator : JC/MD
 Sample : PB137749ZHE#18
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB137749ZHE#18

Quant Time: Jul 20 05:28:39 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	201565	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	334034	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	300543	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	123136	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.971	65	144036	53.169	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 106.340%			
35) Dibromofluoromethane	5.397	113	109459	50.745	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 101.480%			
50) Toluene-d8	8.653	98	401601	49.911	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 99.820%			
62) 4-Bromofluorobenzene	11.085	95	139524	44.916	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 89.840%			
Target Compounds					Qvalue	
16) Acetone	2.386	43	18784	17.526	ug/l #	83
20) Methylene Chloride	2.794	84	14984	6.493	ug/l	98
43) Isopropyl Acetate	6.349	43	28500	5.275	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071921\
Data File : VX023353.D
Acq On : 19 Jul 2021 19:59
Operator : JC/MD
Sample : PB137749ZHE#18
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB137749ZHE#18

Quant Time: Jul 20 05:28:39 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

