

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071921\
 Data File : VX023332.D
 Acq On : 19 Jul 2021 10:24
 Operator : JC/MD
 Sample : VX0719MBL01
 Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0719MBL01

Quant Time: Jul 20 05:22:12 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	217205	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	351308	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	295762	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	117794	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	148623	50.912	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.820%
35) Dibromofluoromethane	5.391	113	114938	50.665	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.320%
50) Toluene-d8	8.653	98	413346	48.845	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.680%
62) 4-Bromofluorobenzene	11.085	95	140581	43.031	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.060%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071921\
Data File : VX023332.D
Acq On : 19 Jul 2021 10:24
Operator : JC/MD
Sample : VX0719MBL01
Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0719MBL01

Quant Time: Jul 20 05:22:12 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.562 min Scan# 734
Delta R.T. -0.000 min
Lab File: VX023332.D
Acq: 19 Jul 2021 10:24




Instrument :
MSVOA_X
ClientSampleId :
VX0719MBL01


Tgt Ion:168 Resp: 217205
Ion Ratio Lower Upper
168 100
99 59.1 45.8 68.6



#33
1,2-Dichloroethane-d4
Concen: 50.912 ug/l
RT: 5.964 min Scan# 800
Delta R.T. 0.000 min
Lab File: VX023332.D
Acq: 19 Jul 2021 10:24

Tgt Ion: 65 Resp: 148623
Ion Ratio Lower Upper
65 100
67 50.0 0.0 105.4









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.769 min Scan# 932
Delta R.T. 0.000 min
Lab File: VX023332.D
Acq: 19 Jul 2021 10:24




Instrument :
MSVOA_X
ClientSampleId :
VX0719MBL01


Tgt Ion:114 Resp: 351308
Ion Ratio Lower Upper
114 100
63 20.2 0.0 40.8
88 15.9 0.0 32.8



#35
Dibromofluoromethane
Concen: 50.665 ug/l
RT: 5.391 min Scan# 706
Delta R.T. -0.006 min
Lab File: VX023332.D
Acq: 19 Jul 2021 10:24

Tgt Ion:113 Resp: 114938
Ion Ratio Lower Upper
113 100
111 102.6 81.9 122.9
192 19.7 16.7 25.1









#50
Toluene-d8
Concen: 48.845 ug/l
RT: 8.653 min Scan# 1241
Delta R.T. -0.000 min
Lab File: VX023332.D
Acq: 19 Jul 2021 10:24




Instrument :
MSVOA_X
ClientSampleId :
VX0719MBL01


Tgt Ion: 98 Resp: 413346
Ion Ratio Lower Upper
98 100
100 64.5 51.9 77.9



#62
4-Bromofluorobenzene
Concen: 43.031 ug/l
RT: 11.085 min Scan# 1640
Delta R.T. 0.000 min
Lab File: VX023332.D
Acq: 19 Jul 2021 10:24

Tgt Ion: 95 Resp: 140581
Ion Ratio Lower Upper
95 100
174 77.1 0.0 154.6
176 74.0 0.0 155.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.006 min
Lab File: VX023332.D
Acq: 19 Jul 2021 10:24

Instrument :
MSVOA_X
ClientSampleId :
VX0719MBL01

Tgt Ion:117 Resp: 295762
Ion Ratio Lower Upper
117 100
82 57.3 47.4 71.2
119 32.1 26.6 39.8



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VX023332.D
Acq: 19 Jul 2021 10:24

Tgt Ion:152 Resp: 117794
Ion Ratio Lower Upper
152 100
115 60.7 41.1 123.3
150 155.6 0.0 349.0

