

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071922\
 Data File : VX030117.D
 Acq On : 19 Jul 2022 15:20
 Operator : JC/MD
 Sample : VX0719MBL01
 Misc : 5.00g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VBLK664

Quant Time: Jul 20 08:37:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML071922WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Jul 20 07:48:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.763	114	270908	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	239047	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	114706	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.367	65	105524	49.938	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	99.880%	
7) Chloroethane-d5	1.666	69	80928	55.343	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	110.680%	
11) 1,1-Dichloroethene-d2	2.300	63	166370	40.704	ug/L	-0.01
Spiked Amount	50.000	Range 60 - 125	Recovery	=	81.400%	
21) 2-Butanone-d5	4.458	46	231609	111.452	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	111.450%	
24) Chloroform-d	5.062	84	217876	50.785	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	101.580%	
26) 1,2-Dichloroethane-d4	5.958	65	136130	50.721	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	101.440%	
32) Benzene-d6	5.976	84	479299	54.471	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	108.940%	
36) 1,2-Dichloropropane-d6	7.311	67	151771	53.786	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	107.580%	
41) Toluene-d8	8.653	98	421954	52.556	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	105.120%	
43) trans-1,3-Dichloroprop...	8.951	79	62254	51.951	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	103.900%	
47) 2-Hexanone-d5	9.384	63	161732	114.297	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	114.300%	
56) 1,1,2,2-Tetrachloroeth...	11.195	84	200711	53.305	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	106.600%	
66) 1,2-Dichlorobenzene-d4	12.323	152	146023	54.917	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	109.840%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071922\
Data File : VX030117.D
Acq On : 19 Jul 2022 15:20
Operator : JC/MD
Sample : VX0719MBL01
Misc : 5.00g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VBLK664

Quant Time: Jul 20 08:37:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXLM071922WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Jul 20 07:48:58 2022
Response via : Initial Calibration

