

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072018\
 Data File : VX003508.D
 Acq On : 20 Jul 2018 14:46
 Operator : JC/SP
 Sample : J3620-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CB-TCLP-11-3-0718

Quant Time: Jul 20 16:31:24 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071918W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 19 17:09:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	279211	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	427931	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	402411	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	223035	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	174755	52.87	ug/l	0.00
Spiked Amount			Recovery	=	105.74%	
35) Dibromofluoromethane	5.51	113	156701	48.08	ug/l	0.00
Spiked Amount			Recovery	=	96.16%	
50) Toluene-d8	8.72	98	598352	48.84	ug/l	0.00
Spiked Amount			Recovery	=	97.68%	
62) 4-Bromofluorobenzene	11.14	95	206827	46.26	ug/l	0.00
Spiked Amount			Recovery	=	92.52%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
8) Diethyl Ether	2.19	74	2397	1.18	ug/l	100
16) Acetone	2.45	43	18670	14.95	ug/l #	89
18) Methyl Acetate	2.78	43	6005	1.53	ug/l #	89
20) Methylene Chloride	2.86	84	406852	134.75	ug/l	91
25) 2-Butanone	4.71	43	4571	2.18	ug/l #	75
30) Chloroform	5.23	83	3480	0.64	ug/l	91
43) Isopropyl Acetate	6.47	43	129471	18.62	ug/l	94
74) N-amyl acetate	6.47	43	129471	19.77	ug/l #	93
80) 1,3,5-Trimethylbenzene	11.51	105	4138	0.33	ug/l	99
84) 1,2,4-Trimethylbenzene	11.81	105	14702	1.14	ug/l	98
95) Naphthalene	13.83	128	97775	6.32	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX072018\
Data File : VX003508.D
Acq On : 20 Jul 2018 14:46
Operator : JC/SP
Sample : J3620-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CB-TCLP-11-3-0718

Quant Time: Jul 20 16:31:24 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071918W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 19 17:09:44 2018
Response via : Initial Calibration

