

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072018\
 Data File : VX003512.D
 Acq On : 20 Jul 2018 16:32
 Operator : JC/SP
 Sample : J4010-01DL 20X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS003RW114R-180713DL

Quant Time: Jul 20 17:10:27 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071918W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 19 17:09:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	276664	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	421273	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	402552	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	222259	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	178663	54.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.10%	
35) Dibromofluoromethane	5.51	113	158400	49.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.74%	
50) Toluene-d8	8.72	98	602673	49.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.94%	
62) 4-Bromofluorobenzene	11.14	95	209066	47.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.00%	
Target Compounds						
65) Chlorobenzene	10.14	112	220118	21.22	ug/l	99
87) 1,3-Dichlorobenzene	12.03	146	28594	3.52	ug/l	98
88) 1,4-Dichlorobenzene	12.10	146	86562	10.34	ug/l	99
91) 1,2-Dichlorobenzene	12.40	146	593477	74.10	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX072018\
Data File : VX003512.D
Acq On : 20 Jul 2018 16:32
Operator : JC/SP
Sample : J4010-01DL 20X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS003RW114R-180713DL

Quant Time: Jul 20 17:10:27 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071918W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 19 17:09:44 2018
Response via : Initial Calibration

