

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX072018\
 Data File : VX003521.D
 Acq On : 20 Jul 2018 21:24
 Operator : JC/SP
 Sample : VX0720WBL03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0720WBL03

Quant Time: Jul 21 03:55:27 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071918W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 19 17:09:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	278586	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	408451	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	381903	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	207377	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	176178	53.42	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.84%	
35) Dibromofluoromethane	5.51	113	154949	49.81	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.62%	
50) Toluene-d8	8.72	98	582847	49.84	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.68%	
62) 4-Bromofluorobenzene	11.14	95	196615	46.07	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.14%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX072018\
Data File : VX003521.D
Acq On : 20 Jul 2018 21:24
Operator : JC/SP
Sample : VX0720WBL03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0720WBL03

Quant Time: Jul 21 03:55:27 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071918W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 19 17:09:44 2018
Response via : Initial Calibration

