

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX072018\
 Data File : VX003529.D
 Acq On : 21 Jul 2018 00:56
 Operator : JC/SP
 Sample : J4014-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE122D1-071218

Quant Time: Jul 27 03:12:22 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071918W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 19 17:09:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	305742	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	427510	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	399466	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	222541	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	183732	50.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.54%	
35) Dibromofluoromethane	5.50	113	174369	53.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.10%	
50) Toluene-d8	8.72	98	607183	49.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.22%	
62) 4-Bromofluorobenzene	11.14	95	209407	46.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.76%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	2.39	101	9346	2.756	ug/l	96
12) 1,1-Dichloroethene	2.38	96	2224	0.712	ug/l	85
30) Chloroform	5.21	83	3560	0.597	ug/l #	83
38) Carbon Tetrachloride	5.79	117	2492	0.550	ug/l #	81
44) Trichloroethene	7.22	130	1185334	285.209	ug/l	98
49) 1,4-Dioxane	7.77	88	1104	13.906	ug/l #	79
64) Tetrachloroethene	9.34	164	1325	0.272	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX072018\
Data File : VX003529.D
Acq On : 21 Jul 2018 00:56
Operator : JC/SP
Sample : J4014-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE122D1-071218

Quant Time: Jul 27 03:12:22 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071918W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 19 17:09:44 2018
Response via : Initial Calibration

