

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX072018\
 Data File : VX003530.D
 Acq On : 21 Jul 2018 01:23
 Operator : JC/SP
 Sample : J4014-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE104D1-071318

Quant Time: Jul 27 03:16:54 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071918W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 19 17:09:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	312883	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	431593	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	398107	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	217117	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	176314	47.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.22%	
35) Dibromofluoromethane	5.50	113	177791	54.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.16%	
50) Toluene-d8	8.72	98	604868	48.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.90%	
62) 4-Bromofluorobenzene	11.14	95	204964	45.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.92%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	2.39	101	5002	1.441	ug/l	97
16) Acetone	2.45	43	7196	5.144	ug/l #	89
44) Trichloroethene	7.21	130	194756	46.418	ug/l	100
64) Tetrachloroethene	9.34	164	2628	0.542	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

