

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX072018\
 Data File : VX003531.D
 Acq On : 21 Jul 2018 01:49
 Operator : JC/SP
 Sample : J4014-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RB01-071218

Quant Time: Jul 27 03:20:39 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071918W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 19 17:09:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	313453	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	428879	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	396724	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	221896	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	180131	48.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.10%	
35) Dibromofluoromethane	5.50	113	180671	55.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.62%	
50) Toluene-d8	8.72	98	607964	49.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.04%	
62) 4-Bromofluorobenzene	11.14	95	207908	46.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.80%	
Target Compounds						
16) Acetone	2.45	43	11061	7.892	ug/l #	82
44) Trichloroethene	7.21	130	3255	0.781	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX072018\
Data File : VX003531.D
Acq On : 21 Jul 2018 01:49
Operator : JC/SP
Sample : J4014-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RB01-071218

Quant Time: Jul 27 03:20:39 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071918W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 19 17:09:44 2018
Response via : Initial Calibration

