

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX072018\
 Data File : VX003532.D
 Acq On : 21 Jul 2018 02:16
 Operator : JC/SP
 Sample : J4014-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE122D3-071218

Quant Time: Jul 23 12:01:21 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071918W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 19 17:09:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	288656	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	422364	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	406745	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	224534	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	187468	54.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.74%	
35) Dibromofluoromethane	5.50	113	172155	53.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.02%	
50) Toluene-d8	8.72	98	611980	50.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.22%	
62) 4-Bromofluorobenzene	11.14	95	209712	47.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.04%	
Target Compounds						
16) Acetone	2.45	43	8131	6.300	ug/l #	81
25) 2-Butanone	4.71	43	9109	4.206	ug/l	95
44) Trichloroethene	7.21	130	36843	8.973	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX072018\
Data File : VX003532.D
Acq On : 21 Jul 2018 02:16
Operator : JC/SP
Sample : J4014-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE122D3-071218

Quant Time: Jul 23 12:01:21 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071918W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 19 17:09:44 2018
Response via : Initial Calibration

