

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX072018\
 Data File : VX003538.D
 Acq On : 21 Jul 2018 04:54
 Operator : JC/SP
 Sample : J4014-16
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE104D2-071318

Quant Time: Jul 23 12:10:11 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071918W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 19 17:09:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	303019	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	426199	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	401359	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	217942	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	184453	51.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.84%	
35) Dibromofluoromethane	5.51	113	179942	55.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.86%	
50) Toluene-d8	8.72	98	603274	49.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.88%	
62) 4-Bromofluorobenzene	11.14	95	204887	46.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.02%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.60	96	88358	23.045	ug/l	89
30) Chloroform	5.21	83	18384	3.113	ug/l	93
42) 1,2-Dichloroethane	6.20	62	4333	0.990	ug/l	95
44) Trichloroethene	7.21	130	260298	62.824	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX072018\
Data File : VX003538.D
Acq On : 21 Jul 2018 04:54
Operator : JC/SP
Sample : J4014-16
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE104D2-071318

Quant Time: Jul 23 12:10:11 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071918W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 19 17:09:44 2018
Response via : Initial Calibration

