

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX072018\
 Data File : VX003539.D
 Acq On : 21 Jul 2018 05:21
 Operator : JC/SP
 Sample : J4014-17
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE105D1-071318

Quant Time: Jul 23 12:13:17 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071918W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 19 17:09:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	297505	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	418080	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	400051	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	220481	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	184886	52.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.00%	
35) Dibromofluoromethane	5.50	113	175176	55.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.02%	
50) Toluene-d8	8.72	98	612818	51.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.40%	
62) 4-Bromofluorobenzene	11.14	95	208177	47.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.32%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	2.39	101	12074	3.659	ug/l	99
16) Acetone	2.45	43	7269	5.465	ug/l #	80
44) Trichloroethene	7.21	130	328563	80.840	ug/l	99
49) 1,4-Dioxane	7.76	88	1343	17.298	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX072018\
Data File : VX003539.D
Acq On : 21 Jul 2018 05:21
Operator : JC/SP
Sample : J4014-17
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE105D1-071318

Quant Time: Jul 23 12:13:17 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071918W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 19 17:09:44 2018
Response via : Initial Calibration

