

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072019\
 Data File : VX010991.D
 Acq On : 19 Jul 2019 10:20
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 20 01:16:13 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 17 06:57:45 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	108	0.00
2 T	Dichlorodifluoromethane	0.467	0.544	-16.5	110	0.00
3 P	Chloromethane	0.460	0.437	5.0	103	0.00
4 C	Vinyl Chloride	0.469	0.476	-1.5#	106	0.00
5 T	Bromomethane	0.264	0.213	19.3	92	0.00
6 T	Chloroethane	0.298	0.279	6.4	101	0.00
7 T	Trichlorofluoromethane	0.794	0.821	-3.4	111	0.00
8 T	Diethyl Ether	0.274	0.267	2.6	109	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.453	0.477	-5.3	116	0.00
10 T	Methyl Iodide	0.595	0.469	21.2#	80	0.00
11 T	Tert butyl alcohol	0.124	0.109	12.1	102	0.00
12 CM	1,1-Dichloroethene	0.445	0.454	-2.0#	109	0.00
13 T	Acrolein	0.072	0.065	9.7	100	0.00
14 T	Allyl chloride	0.647	0.661	-2.2	111	0.00
15 T	Acrylonitrile	0.250	0.240	4.0	106	0.00
16 T	Acetone	0.240	0.257	-7.1	117	0.00
17 T	Carbon Disulfide	1.142	1.126	1.4	105	0.00
18 T	Methyl Acetate	0.547	0.518	5.3	111	0.00
19 T	Methyl tert-butyl Ether	1.434	1.451	-1.2	113	0.00
20 T	Methylene Chloride	0.509	0.495	2.8	110	0.00
21 T	trans-1,2-Dichloroethene	0.473	0.469	0.8	108	0.00
22 T	Diisopropyl ether	1.364	1.376	-0.9	112	0.00
23 T	Vinyl Acetate	1.156	1.212	-4.8	111	0.00
24 P	1,1-Dichloroethane	0.792	0.777	1.9	110	0.00
25 T	2-Butanone	0.345	0.352	-2.0	110	0.00
26 T	2,2-Dichloropropane	0.651	0.695	-6.8	117	0.00
27 T	cis-1,2-Dichloroethene	0.547	0.541	1.1	111	0.00
28 T	Bromochloromethane	0.373	0.386	-3.5	117	0.00
29 T	Tetrahydrofuran	0.218	0.211	3.2	106	0.00
30 C	Chloroform	0.856	0.860	-0.5#	114	0.00
31 T	Cyclohexane	0.722	0.737	-2.1	114	0.00
32 T	1,1,1-Trichloroethane	0.741	0.765	-3.2	112	0.00
33 S	1,2-Dichloroethane-d4	0.532	0.526	1.1	111	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	106	0.00
35 S	Dibromofluoromethane	0.317	0.324	-2.2	109	0.00
36 T	1,1-Dichloropropene	0.411	0.437	-6.3	114	0.00
37 T	Ethyl Acetate	0.421	0.438	-4.0	107	0.00
38 T	Carbon Tetrachloride	0.428	0.470	-9.8	115	0.00
39 T	Methylcyclohexane	0.533	0.574	-7.7	115	0.00
40 TM	Benzene	1.258	1.287	-2.3	109	0.00
41 T	Methacrylonitrile	0.239	0.237	0.8	112	0.00
42 TM	1,2-Dichloroethane	0.437	0.457	-4.6	115	0.00
43 T	Isopropyl Acetate	0.692	0.716	-3.5	110	0.00
44 TM	Trichloroethene	0.372	0.379	-1.9	110	0.00
45 C	1,2-Dichloropropane	0.320	0.329	-2.8#	111	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\WX072019\
 Data File : VX010991.D
 Acq On : 19 Jul 2019 10:20
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 20 01:16:13 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 17 06:57:45 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.229	0.239	-4.4	111	0.00
47 T	Bromodichloromethane	0.407	0.451	-10.8	115	0.00
48 T	Methyl methacrylate	0.337	0.351	-4.2	112	0.00
49 T	1,4-Dioxane	0.009	0.009	0.0	103	0.00
50 S	Toluene-d8	1.190	1.220	-2.5	108	0.00
51 T	4-Methyl-2-Pentanone	0.448	0.468	-4.5	109	0.00
52 CM	Toluene	0.826	0.856	-3.6#	111	0.00
53 T	t-1,3-Dichloropropene	0.438	0.502	-14.6	116	0.00
54 T	cis-1,3-Dichloropropene	0.487	0.535	-9.9	111	0.00
55 T	1,1,2-Trichloroethane	0.340	0.354	-4.1	112	0.00
56 T	Ethyl methacrylate	0.498	0.546	-9.6	112	0.00
57 T	1,3-Dichloropropane	0.540	0.562	-4.1	112	0.00
58 T	2-Chloroethyl Vinyl ether	0.238	0.215	9.7	93	0.00
59 T	2-Hexanone	0.348	0.371	-6.6	108	0.00
60 T	Dibromochloromethane	0.345	0.393	-13.9	114	0.00
61 T	1,2-Dibromoethane	0.361	0.383	-6.1	111	0.00
62 S	4-Bromofluorobenzene	0.438	0.477	-8.9	113	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	107	0.00
64 T	Tetrachloroethene	0.393	0.410	-4.3	116	0.00
65 PM	Chlorobenzene	1.027	1.039	-1.2	113	0.00
66 T	1,1,1,2-Tetrachloroethane	0.361	0.393	-8.9	113	0.00
67 C	Ethyl Benzene	1.748	1.857	-6.2#	114	0.00
68 T	m/p-Xylenes	0.671	0.711	-6.0	114	0.00
69 T	o-Xylene	0.654	0.690	-5.5	113	0.00
70 T	Styrene	1.091	1.183	-8.4	113	0.00
71 P	Bromoform	0.288	0.329	-14.2	116	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	108	0.00
73 T	Isopropylbenzene	3.378	3.547	-5.0	117	0.00
74 T	N-amyl acetate	1.291	1.346	-4.3	110	0.00
75 P	1,1,2,2-Tetrachloroethane	1.106	1.106	0.0	110	0.00
76 T	1,2,3-Trichloropropane	0.902	0.932	-3.3	112	0.00
77 T	Bromobenzene	0.932	0.940	-0.9	113	0.00
78 T	n-propylbenzene	3.733	4.004	-7.3	117	0.00
79 T	2-Chlorotoluene	2.242	2.333	-4.1	115	0.00
80 T	1,3,5-Trimethylbenzene	2.823	2.987	-5.8	115	0.00
81 T	trans-1,4-Dichloro-2-butene	0.328	0.347	-5.8	113	0.00
82 T	4-Chlorotoluene	2.610	2.755	-5.6	117	0.00
83 T	tert-Butylbenzene	2.773	2.935	-5.8	115	0.00
84 T	1,2,4-Trimethylbenzene	2.845	3.021	-6.2	115	0.00
85 T	sec-Butylbenzene	3.249	3.531	-8.7	117	0.00
86 T	p-Isopropyltoluene	3.004	3.298	-9.8	118	0.00
87 T	1,3-Dichlorobenzene	1.617	1.701	-5.2	116	0.00
88 T	1,4-Dichlorobenzene	1.659	1.698	-2.4	114	0.00
89 T	n-Butylbenzene	2.570	2.930	-14.0	120	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072019\
Data File : VX010991.D
Acq On : 19 Jul 2019 10:20
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jul 20 01:16:13 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 17 06:57:45 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.474	0.527	-11.2	118	0.00
91 T	1,2-Dichlorobenzene	1.603	1.638	-2.2	112	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.231	0.244	-5.6	112	0.00
93 T	1,2,4-Trichlorobenzene	1.086	1.191	-9.7	118	0.00
94 T	Hexachlorobutadiene	0.538	0.591	-9.9	121	0.00
95 T	Naphthalene	3.208	3.437	-7.1	112	0.00
96 T	1,2,3-Trichlorobenzene	1.081	1.170	-8.2	117	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6