

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072020\
 Data File : VX017435.D
 Acq On : 20 Jul 2020 15:42
 Operator : JC/SP
 Sample : L3329-11
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ERM-16R

Quant Time: Jul 21 02:29:38 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071320W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 14 02:59:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	240542	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	422362	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	439443	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	241675	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	166001	54.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.98%	
35) Dibromofluoromethane	5.47	113	145051	50.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.96%	
50) Toluene-d8	8.70	98	522026	50.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.74%	
62) 4-Bromofluorobenzene	11.12	95	222079	54.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.96%	
Target Compounds						
16) Acetone	2.42	43	6665	6.188	ug/l	89
39) Methylcyclohexane	7.43	83	1917	0.454	ug/l	88
83) tert-Butylbenzene	11.76	119	3138	0.252	ug/l	77
85) sec-Butylbenzene	11.93	105	5131	0.353	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX072020\
Data File : VX017435.D
Acq On : 20 Jul 2020 15:42
Operator : JC/SP
Sample : L3329-11
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ERM-16R

Quant Time: Jul 21 02:29:38 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071320W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 14 02:59:51 2020
Response via : Initial Calibration

