

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072020\
 Data File : VX017438.D
 Acq On : 20 Jul 2020 16:50
 Operator : JC/SP
 Sample : PB130293TB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB130293TB

Quant Time: Jul 21 02:37:51 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071320W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 14 02:59:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	266300	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	462884	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	489285	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	277328	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	175639	52.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.10%	
35) Dibromofluoromethane	5.46	113	155042	49.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.44%	
50) Toluene-d8	8.70	98	576236	51.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.46%	
62) 4-Bromofluorobenzene	11.12	95	246552	55.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.38%	
Target Compounds						
16) Acetone	2.42	43	6269	5.258	ug/l	99
20) Methylene Chloride	2.84	84	17057	5.863	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX072020\
Data File : VX017438.D
Acq On : 20 Jul 2020 16:50
Operator : JC/SP
Sample : PB130293TB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB130293TB

Quant Time: Jul 21 02:37:51 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071320W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 14 02:59:51 2020
Response via : Initial Calibration

